

	A		
	Application No.	Applicant(s)	
	10/774,450	TONAMI ET AL.	
Notice of Allowability •	Examiner	Art Unit	
	Djura Malevic	2884	
The MAILING DATE of this communication appear All claims being allowable, PROSECUTION ON THE MERITS IS (Cherewith (or previously mailed), a Notice of Allowance (PTOL-85) or NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIG of the Office or upon petition by the applicant. See 37 CFR 1.313 a	OR REMAINS) CLOSED rother appropriate commitments. This application is	in this application. If not included munication will be mailed in due cours	se. <b>THIS</b>
1. This communication is responsive to <u>2/10/2004</u> .	•		
2. The allowed claim(s) is/are <u>1-10</u> .			
<ul> <li>3.  Acknowledgment is made of a claim for foreign priority under a)  All b)  Some* c)  None of the:</li> <li>1.  Certified copies of the priority documents have b</li> </ul>		) or (f).	
2.  Certified copies of the priority documents have b	•	ion No.	
3. Copies of the certified copies of the priority docu	• •		from the
International Bureau (PCT Rule 17.2(a)).			
* Certified copies not received:	•		
Applicant has THREE MONTHS FROM THE "MAILING DATE" of noted below. Failure to timely comply will result in ABANDONME THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.		le a reply complying with the require	ments
4. A SUBSTITUTE OATH OR DECLARATION must be submitted informal patent application (PTO-152) which gives			CE OF
5. CORRECTED DRAWINGS (as "replacement sheets") must be	be submitted.		
(a) I including changes required by the Notice of Draftspersor	n's Patent Drawing Revie	ew ( PTO-948) attached	
1)  hereto or 2)  to Paper No./Mail Date	•		
(b) ☐ including changes required by the attached Examiner's A Paper No./Mail Date	Amendment / Comment	or in the Office action of	
Identifying indicia such as the application number (see 37 CFR 1.84 each sheet. Replacement sheet(s) should be labeled as such in the			k) of
6. DEPOSIT OF and/or INFORMATION about the deposit attached Examiner's comment regarding REQUIREMENT FO	OF BIOLOGICAL MATERIES OF B	TERIAL must be submitted. Note IOLOGICAL MATERIAL.	the
Attachment(s) 1. ⊠ Notice of References Cited (PTO-892)	5 🗆 Notice of	Informal Patent Application (PTO-15)	2)
2. ☐ Notice of Draftperson's Patent Drawing Review (PTO-948)		Summary (PTO-413),	۷)
	Paper No	o./Mail Date	
<ol> <li>Information Disclosure Statements (PTO-1449 or PTO/SB/08) Paper No./Mail Date <u>2/10/04</u></li> </ol>	), 7. ☐ Examiner	's Amendment/Comment	
Examiner's Comment Regarding Requirement for Deposit of Biological Material	8. 🛭 Examiner	's Statement of Reasons for Allowand	се
	9. 🗌 Other	<u></u> ·	
	_		
•	•		

Application/Control Number: 10/774,450

Art Unit: 2884

## **DETAILED ACTION**

## **Information Disclosure Statement**

The prior art cited in the information disclosure statement (IDS) submitted on 02/10/2004 has been considered.

## Allowable Subject Matter

Claims 1 -10 are allowed.

The following is an examiner's statement of reasons for allowance:

With regards to claims 1 and 6, the prior art of record does not disclose or suggest a radiation detector comprising a light guide formed of a lattice frame member that is integrally formed with a cured liquid resin wherein, the lattice frame member forms partition walls in the cured liquid resin to provide compartments. While aspects of the prior art do address radiation detectors wherein, the light guide is disposed between the scintillators and the photoelectron multipliers, the prior art is further silent with regards to a lattice frame integrally formed with the cured liquid resin and constructing partition walls once the liquid resin sets.

References, such as Cherry et al. (US Patent 6,552,348 B2) discloses a light guide that was cut into individually tapered light guides, wherein the individual cut light guides would match the scintillation crystal array and the active face of the PMT, teaching away from the applicant's discloser. Cherry fails to suggest any modifications to include a lattice frame integrally formed with the cured liquid resin.

Application/Control Number: 10/774,450

Art Unit: 2884

As such, applicant's discloser provides a novel and nonobvious improvement over the prior art. Accordingly, claims 2-5 and 7-10 are allowable because they further limit claims 1 and 6.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

## Conclusion

The prior art made of record and not relied upon is considered pertinent to applicant's disclosure. Saitoh *et al.* (US Pub. 20030099118 A1) discloses a light guide made by cutting acrylic resins.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Djura Malevic whose telephone number is 571.272.5975. The examiner can normally be reached on Monday - Friday between 8:30am and 4:00pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, David Porta can be reached on (571) 272-2444. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Application/Control Number: 10/774,450

Art Unit: 2884

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <a href="http://pair-direct.uspto.gov">http://pair-direct.uspto.gov</a>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

Djura Malevic
Patent Examiner
Art Unit 2884
571.272.5975

DAMD PORTA
SUPERVISORY PATENT EXAMINEF
TECHNOLOGY CENTER 2800